

File : Q:\GCMS-08 Minimal\DATA\20190620\BSB\19-.D
Operator : JP/KM
Acquired : 20 Jun 2019 21:02 using AcqMethod KERLA-2Split-2uL-24min.M
Instrument : GCMS 8
Sample Name: ENISO9377-2-3
Misc Info : Resample in Hexane
Vial Number: 48



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

