

File : D:\MassHunter\GCMS\1\data\02142023\20- ALR.D
Operator :
Acquired : 14 Feb 2023 16:47 using AcqMethod Hy SV DB-5MS UI20M 0.18 2 split FAST .M
Instrument : GCMS 9
Sample Name : F-100S
Misc Info : RT: Dibenzofuran in Isooctane
Vial Number: 1



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

