

File :D:\MassHunter\GCMS\1\data\11221901\BSB\0.D  
Operator : KM  
Acquired : 22 Nov 2019 10:05 using AcqMethod CHICK\_2019.M  
Instrument : GCMS 6  
Sample Name: F-4025  
Misc Info : RT: 2,3,7,8-Tetrachlorodibenzofuran in Toluene  
Vial Number: 19



AccuStandard®

Column: DB-5, 30m, 0.25 ID, 0.25um  
Oven Program: 30c, 2m, 30c/min to 320c, final 2min.  
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow  
Split 20:1, 1ul inj.; GC/MS; INJ 280c.

OK

