

File : L:\QC Data\GCMS-07 Minimal\DATA\05122301\38 - SAK.D
Operator : Organic QC department
Acquired : 13 May 2023 03:40 using AcqMethod KERLA-2017_Short_15Min_S100.M
Instrument : 5977B MSD
Sample Name: FAEE-004S
Misc Info :
Vial Number: 114



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 100:1, 1µl inj.; GC/MS; Injector temp 270°C

