

File : Q:\GCMS-06\DATA\11081801\BSB\16-
Operator : D.C
Acquired : 08 Nov 2018 16:55 using AcqMethod CHICK_2017_VOC-SV.M
Instrument : GCMS 6
Sample Name: FRS-011S
Misc Info : FR-1138 (Dow) in Toluene
Vial Number: 50



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 35c, 17c/min to 340, 8min
GC Parameters: Cons. Split, 120mL/min constant flow
Split 100:1, 1uL inj.; GC/MS; INJ 200c

