


File :D:\MassHunter\GCMS\1\DATA\06202201\34- KD.D
Operator : Organic QC department
Acquired : 20 Jun 2022 18:27 using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name: FRS-017S (222061362)
Misc Info : Hexachlorobutadiene in Toluene
Vial Number: 64

 **AccuStandard**[®]
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

