

File : Q:\GCMS-05 Minimal\DATA\12031901\BSB\01-.D
Operator : [BSB2]D.C
Acquired : 3 Dec 2019 13:59 using AcqMethod BDE_SPLITLESS.M
Instrument : Jack MSD
Sample Name: FRS-034S
Misc Info : FRS-034 in Toluene
Vial Number: 3

Column: DB-5MS, 15m, 0.25 ID, 0.25um
Oven Program: 120c, 20c/min to 340c, 5min
GC Parameters: Cons. Split, 12psi constant flow
Splitless, 1uL inj.; GC/MS; INJ 270c

