

File :D:\MassHunter\GCMS\1\data\20220607\06 - SAK.D
Operator : Organic QC Lab
Acquired : 07 Jun 2022 10:07 using AcqMethod Hy_VOC-SV_15cm_S50.M
Instrument : GCMS 8
Sample Name: FU-012
Misc Info :
Vial Number: 107



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

