

File : Q:\GCMS-06 Minimal\DATA\060821\BSB\21- KM.D  
Operator : Organic QC Lab  
Acquired : 09 Jun 2021 01:29 using AcqMethod CHICK\_2019\_S100.M  
Instrument : GCMS 6  
Sample Name: FU-020-D-40X  
Misc Info : Hydraulic fluid in MeCl2  
Vial Number: 148



**AccuStandard®**

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

