

Sample Name : GBS-008S

Acq. Operator : SYSTEM

Seq. Line : 2

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2

Location : P2-F-02

Injection Date : 6/14/2024 6:39:23 AM

Inj : 2

Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.500 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\061424\061424 2024-06-14 06-21-36
\DNP.H.M

Last changed : 12/4/2019 12:27:07 PM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 5/30/2024 7:55:21 AM by SYSTEM

(modified after loading)

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

