

File :Q:\GCMS-08 Minimal\DATA\20211027\50- (JP).D
Operator : Organic QC Lab
Acquired : 28 Oct 2021 03:25 using AcqMethod Hy_Pest_15cm_S2_3ulinj.M
Instrument : GCMS 8
Sample Name: GPC-CC-B-WL-10ML
Misc Info : Aroclor 1016 + 1260; in dcm
Vial Number: 65



Column: DB-5, 30m, 0.25ID, 0.25um
Oven: 80c, 2m, 30c/m to 320, final 2min
GC para: Pulse split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj, GCMS, Inj 280c

