

File :D:\MassHunter\GCMS\1\data\06012022\12- ALR.D  
Operator : ALR  
Acquired : 01 Jun 2022 14:51 using AcqMethod Hy SV DB-5MS UI20M 0.18 50 split FAST .M  
Instrument : GCMS 9  
Sample Name: H-152S  
Misc Info : Naphthalene in Toluene  
Vial Number: 35



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

