

File :Q:\GCMS-06 Minimal\DATA\072720\18- KM.D
Operator : Organic QC Lab
Acquired : 27 Jul 2020 17:39 using AcqMethod CHICK_2017_13MIN.M
Instrument : GCMS 6
Sample Name: H-162S
Misc Info : 1-Methylphenanthrene in Toluene
Vial Number: 5



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

