

File :Q:\GCMS-08 Minimal\DATA\20220214\13- =JP=.D
Operator : Organic QC Lab
Acquired : 14 Feb 22 02:03 pm using AcqMethod Hy_Pest_24cm_S1.M
Instrument : GCMS 8
Sample Name: H-247S
Misc Info : Dibenz(a,e)fluoranthene; in toluene
Vial Number: 56



Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

