

File : L:\QC Data\GCMS-09 Minimal\DATA\09192022\21 -SAK.D  
 Operator :  
 Acquired : 19 Sep 2022 14:47 using AcqMethod Hy SV DB-5MS UI20M 0.18 40 split TIME17 MIN .M  
 Instrument : GCMS 9  
 Sample Name: S-25229-10X  
 Misc Info : Custom PAH std in Tol  
 Vial Number: 101

Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm  
 Oven Program: 45°C, 13°C/min to 290°C (hold 5min)  
 GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
 Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

