

```
=====
Acq. Operator   : SYSTEM                               Seq. Line :    2
Acq. Instrument : LCMS-1                               Location  : P1-A1
Injection Date  : 10/30/2024 10:47:26 AM              Inj       :    2
                                                    Inj Volume: 1.000 µl
```

Different Inj Volume from Sample Entry! Actual Inj Volume : 5.000 µl

```
Acq. Method     : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 1/5/2023 12:01:53 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed    : 10/23/2024 9:11:26 AM by SYSTEM
                 (modified after loading)
```

```
Method Info     : Operator: Organic QC lab
                 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                 Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                 :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                 particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
                 MS Detector. No UV Reference.
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

