



```

=====
Acq. Operator   : SYSTEM                      Seq. Line :   13
Acq. Instrument : LCMS-1                     Location  :   P2-E1
Injection Date  : 12/11/2019 6:10:46 PM      Inj       :    1
                                           Inj Volume: 1.000 µl
Different Inj Volume from Sample Entry! Actual Inj Volume : 2.000 µl

```

```

Acq. Method    : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 10/18/2019 4:18:14 PM by SYSTEM
Analysis Method: C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 12/11/2019 1:10:22 PM by SYSTEM
                (modified after loading)

```

```

Method Info    : Operator: GA
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml
                /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.
                Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV
                Reference.

```

Sample Info : 6"-Hydroxy-2,2',2",4-tetrabromodiphenoxybenzene @50ug/mL in AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

