

File :D:\MassHunter\GCMS\1\data\20221219\06- _JP_.D
Operator : Organic QC Lab
Acquired : 19 Dec 2022 09:08 using AcqMethod Hy_Pest_24cm_S1.M
Instrument : GCMS 8
Sample Name: HPCB-4016S
Misc Info : 4,6'-hydroxy-22'4'6-tetraPCB; in iso
Vial Number: 13



Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

