

File :Q:\GCMS-06 Minimal\DATA\101320\27- KM.D
Operator : Organic QC Lab
Acquired : 13 Oct 2020 22:07 using AcqMethod CHICK_2019_VOC-SV_S100.M
Instrument : GCMS 6
Sample Name: HS-001S-40X
Misc Info : Lacquer Thinner in MeOH
Vial Number: 3



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

