

File :B:\QC Data\GCMS-04 Minimal\DATA\090619\09-.D
Operator : RFXM / ZA / DC
Acquired : 6 Sep 2019 14:50 using AcqMethod VOC-200SPLIT HOT.M
Instrument : GCMS4
Sample Name: HS-004S-D-40X
Misc Info : TURPENTINE
Vial Number: 94



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

