


.D
File : C:\msdchem\1\data\01212001\BSB\04-
Operator : [BSB1]Organic QC Department
Acquired : 21 Jan 2020 13:32 using AcqMethod DECKER2014.M
Instrument : Jack MSD
Sample Name: HXBCD-01
Misc Info : a-Hexabromocyclododecane in Toluene
Vial Number: 6

 **AccuStandard**
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

