

File :Q:\GCMS-08\DATA\20190304\04-  
Operator : JP/KM  
Acquired : 04 Mar 2019 16:49 using AcqMethod KERLA-2017-100uL draw speed.M  
Instrument : GCMS 8  
Sample Name: L-101S-CN  
Misc Info : 2,4-Dimethylaniline in AcCN  
Vial Number: 2



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

