

File :D:\MassHunter\GCMS\1\DATA\02202301\20- -KD.D  
Operator : Organic QC department  
Acquired : 20 Feb 23 02:23 pm using AcqMethod KERLA-2017.M  
Instrument : 5977B MSD  
Sample Name: M-005A  
Misc Info :  
Vial Number: 72



Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

