



Acq. Operator : SYSTEM Seq. Line : 4  
 Sample Operator : SYSTEM  
 Acq. Instrument : LCMS-2 Location : P1-E-02  
 Injection Date : 9/19/2022 9:31:13 AM Inj : 4  
 Inj Volume : 2.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\09192201\09192201 2022-09-19 08-31-51  
 \PFAA.M

Last changed : 1/26/2021 10:30:23 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M

Last changed : 3/23/2022 11:04:54 AM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

