



Acq. Operator : SYSTEM Seq. Line : 40
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2 Location : P1-A-11
Injection Date : 6/7/2022 6:57:54 PM Inj : 2
Inj Volume : 2.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\06072201\06072201 2022-06-07 06-39-34\PF AA.M
Last changed : 1/26/2021 10:30:23 AM by SYSTEM
Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M
Last changed : 5/16/2022 12:35:31 PM by SYSTEM (modified after loading)

Method Info : OPERATOR: Organic QC lab
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

