


File : L:\QC Data\GCMS-06 Minimal\DATA\022820\14-2D
Operator : Organic QC Lab
Acquired : 28 Feb 2020 14:01 using AcqMethod M502-FAST.M
Instrument : GCMS 6
Sample Name: M-1667A-03
Misc Info :
Vial Number: 122

 **AccuStandard®**
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

