Print of window 38: Current Chromatogram(s)
Data File : D:\DATA\11302023\35- KM.D
Sample Name : M-1667A-DERV-10ML (223111477)

==========

Acq. Operator : SYSTEM
Seq. Line : 35
Acq. Instrument : LCMS-1
Location : P2-B6
Injection Date : 11/30/2023 5:01:21 PM
Inj : 1
Inj Volume : 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 1/5/2023 12:01:53 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 11/29/2023 8:30:46 AM by SYSTEM
(modified after loading)

Method Info : Operator: Organic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml /min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Sample Info : 2,4-Dinitrophenylhydrazine in ACN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)