



=====

|                                       |                       |
|---------------------------------------|-----------------------|
| Acq. Operator : SYSTEM                | Seq. Line : 1         |
| Sample Operator : SYSTEM              |                       |
| Acq. Instrument : LCMS-2              | Location : P1-B-06    |
| Injection Date : 6/22/2020 3:48:50 PM | Inj : 1               |
|                                       | Inj Volume : 1.000 µl |

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.100 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\06222001\06222001 2020-06-22 15-47-52  
\DNPH.M

Last changed : 12/4/2019 12:27:07 PM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M

Last changed : 6/22/2020 4:07:02 PM by SYSTEM  
(modified after loading)

Method Info : OPERATOR: Organic QC lab  
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

