

File :Q:\GCMS-04 Minimal\DATA\012722\57- YX.D
Operator : Organic QC Lab
Acquired : 28 Jan 2022 15:14 using AcqMethod VOC-100SPLIT DECKER.M
Instrument : GCMS4
Sample Name: M-502-03
Misc Info : Bromochloromethane in MeOH
Vial Number: 29



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

