

File : Q:\GCMS-08 Minimal\DATA\20230608\32- JP\_.D  
Operator : Organic QC Lab  
Acquired : 08 Jun 2023 15:22 using AcqMethod Hy\_Pest\_24cm\_S2\_2ulinj.M  
Instrument : GCMS 8  
Sample Name: M-505-ASL  
Misc Info : in acetone  
Vial Number: 37



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

