

File :Q:\GCMS-08 Minimal\DATA\20230731\39 -(JP).D
Operator : Organic QC Lab
Acquired : 01 Aug 2023 02:31 using AcqMethod Hy_M507QC_15cm_Puls40_Spltls_5ul_mtbe.M
Instrument : GCMS 8
Sample Name: M-507-QC
Misc Info :in mtbe
Vial Number: 15



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5m in, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

