

File :Q:\GCMS-08 Minimal\DATA\20191025\BSB\16- .D  
Operator : KM / JP / ZA  
Acquired : 25 Oct 2019 21:34 using AcqMethod KERLA-2017-100uL draw speed.M  
Instrument : GCMS 8  
Sample Name: M-521  
Misc Info : EPA Method 521 Analyte Stock in MeCl2  
Vial Number: 34



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

