

File : Q:\GCMS-06\DATA\07111801\06-D
 Operator : D.C
 Acquired : 11 Jul 18 11:31 am using AcqMethod CHICK_2017_Plus 200 EMV.M
 Instrument : GCMS 6
 Sample Name: M-525-REG-EA
 Misc Info : Regulated SV in Ethyl acetate
 Vial Number: 55



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
 Oven Program: 80c, 17c/min to 340c, 8min
 GC Parameters: Cons. Split, 12psi constant flow
 Split 20:1, 1uL inj.; GC/MS; INJ 270c

