

File : Q:\GCMS-07\DATA\07191801\09-
Operator : K.D
Acquired : 19 Jul 2018 12:41 using AcqMethod KERLA-2017.M
Instrument : 5977B MSD
Sample Name: M-525.3-06
Misc Info:
Vial Number: 75

Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

