Print of window 38: Current Chromatogram(s)
Data File : D:\DATA\03192024\25- KM.D
Sample Name : M-531-03

Acq. Operator	: SYSTEM	Seq. Line	: 6
Acq. Instrument	: LCMS-1	Location	: P2-A1
Injection Date	: 3/19/2024 4:18:42 PM	Inj	: 1
		Inj Volume	: 1.000 µl

Acq. Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 1/5/2023 12:01:53 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed : 3/20/2024 7:51:53 AM by SYSTEM
(modified after loading)
Method Info : Operator: Organic QC lab
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV Reference.

Sample Info : Oxamyl in ACN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)