

=====

Acq. Operator : SYSTEM Seq. Line : 10

Sample Operator : SYSTEM

Acq. Instrument : LCMS-2 Location : P2-D-03

Injection Date : 7/9/2020 5:56:50 PM Inj : 1

Inj Volume : 1.000 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\07092001\07092001 2020-07-09 15-05-43  
\Purity\_195-315.M

Last changed : 5/8/2020 10:12:19 AM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity\_195-315.M

Last changed : 7/10/2020 11:05:06 AM by SYSTEM

(modified after loading)

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

