



=====  
Acq. Operator : SYSTEM Seq. Line : 27  
Acq. Instrument : LCMS-1 Location : P2-B2  
Injection Date : 6/26/2019 9:37:05 PM Inj : 1  
Inj Volume : 1.000 µl

Acq. Method : D:\DATA\06261901\06261903 2019-06-26 13-32-08\Purity195\_315.M  
Last changed : 6/26/2019 1:32:08 PM by SYSTEM  
Analysis Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
Last changed : 6/26/2019 1:55:40 PM by SYSTEM  
(modified after loading)

Method Info : Operator: GA  
Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %  
Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate:0.3 ml  
/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size.  
Agilent 1290 Infinity II LC system, Agilent 6120 Quadrople MS Detector. No UV  
Reference.

Sample Info : 3-Hydrocarbofuran

Current Chromatogram(s)

