

```
=====
Acq. Operator   : SYSTEM                               Seq. Line : 12
Acq. Instrument : LCMS-1                             Location  : P1-B8
Injection Date  : 12/26/2023 4:25:10 PM             Inj       : 1
                                                    Inj Volume : 1.000 µl
```

```
Acq. Method    : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 1/5/2023 12:01:53 PM by SYSTEM
Analysis Method : C:\CHEM32\1\METHODS\Purity195_315.M
Last changed   : 11/29/2023 8:30:46 AM by SYSTEM
```

```
(modified after loading)
Method Info    : Operator: Organic QC lab
                Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %
                Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate
                :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron
                particle size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole
                MS Detector. No UV Reference.
```

Sample Info : Baygon in AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

