

```

Acq. Operator   : SYSTEM                      Seq. Line :   2
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                     Location  :   P1-C-09
Injection Date  : 2/14/2020 11:57:42 AM      Inj       :   1
                                           Inj Volume: 1.000 µl

Acq. Method     : C:\Users\Public\Documents\ChemStation\1\Data\02142001\01202001 2020-02-14
                   11-37-48\Purity_195-315.M
Last changed    : 12/3/2019 1:25:24 PM by SYSTEM
Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
Last changed    : 2/12/2020 2:40:50 PM by SYSTEM
                   (modified after loading)

Method Info     : General gradient method starting @100% water for 1 min @0.3 ml/min,the 100%
                   Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax
                   Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity
                   II LC/MS system, Agilent 6125 quadrupole MS detector.
    
```

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

