



**AccuStandard**  
Leader in Analytical Reference Standards

```

=====
Acq. Operator   : SYSTEM                               Seq. Line   :    2
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2                                   Location    : P1-F-07
Injection Date  : 5/31/2022 3:14:23 PM             Inj         :    2
                                                    Inj Volume  : 1.000 µl
  
```

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\05312201\05312201 2022-05-31 14-54-27\Purity\_195-315.M

Last changed : 3/23/2022 11:04:54 AM by SYSTEM

Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity\_195-315.M

Last changed : 5/16/2022 12:35:31 PM by SYSTEM  
(modified after loading)

Method Info : OPERATOR: Organic QC lab  
General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : Methiocarb in AcCN

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

