

File :Q:\GCMS-04\DATA\081418\21-D
Operator : KM
Acquired : 14 Aug 2018 23:52 using AcqMethod VOC-100SPLIT DECKER.M
Instrument : GCMS4
Sample Name: M-541-IS
Misc Info : EO
Vial Number: 77



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

