



```

=====
Acq. Operator   : SYSTEM                               Seq. Line :    2
Sample Operator : SYSTEM                               Location  :   P2-F3
Acq. Instrument : LCMS-2                              Inj       :    2
Injection Date  : 12/23/2024 9:17:16 AM              Inj Volume: 1.000 µl
Difference Inj Volume from Sample Entry! Actual Inj Volume : 2.000 µl
    
```

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\122324\122324 2024-12-23 08-57-25 \Purity_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Current Chromatogram(s)

