

File :D:\MassHunter\GCMS\1\data\05262023\57- ALR.D
Operator :
Acquired : 27 May 2023 01:08 using AcqMethod Hy SV DB-5MS UI20M 0.18 2 split FAST .M
Instrument : GCMS 9
Sample Name: M-548B
Misc Info : Endothall in Water
Vial Number: 27



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

