


File :Q:\GCMS-08 Minimal\DATA\20220114\09- (JP).D
Operator : Organic QC Lab
Acquired : 14 Jan 22 12:40 pm using AcqMethod Hy_Pest_24cm_S1_2ulinj.M
Instrument : GCMS 8
Sample Name: M-551.1C
Misc Info :in acetone
Vial Number: 54

 **AccuStandard**
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

