


File : L:\QC Data\GCMS-06 Minimal\DATA\071221\23- ZA.D
Operator : Organic QC Lab
Acquired : 12 Jul 2021 17:34 using AcqMethod CHICK_2019_VOC-SV.M
Instrument : GCMS 6
Sample Name: M-552-R-02
Misc Info :
Vial Number: 51

 **AccuStandard®**
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 μ m
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1 μ L inj.; GC/MS; INJ 270c

