

File : L:\QC Data\GCMS-09 Minimal\DATA\10192022\14 -SAK.D
Operator :
Acquired : 19 Oct 2022 13:05 using AcqMethod ALR 19 min 100 split FAST .M
Instrument : GCMS 9
Sample Name : M-552.1-IS
Misc Info :
Vial Number : 118



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

