

File :Q:\GCMS-06 Minimal\DATA\022322\42- YX.D
Operator : Organic QC Lab
Acquired : 24 Feb 2022 01:02 using AcqMethod CHICK_2019_VOC-SV_2S_2uL.M
Instrument : GCMS 6
Sample Name: M-552.2A-05
Misc Info : Dibromoacetic acid
Vial Number: 119



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

