

File :Q:\GCMS-06 Minimal\DATA\030122\28-YX.D
Operator : Organic QC Lab
Acquired : 01 Mar 2022 17:46 using AcqMethod CHICK_2019_VOC-SV_2S_2uL.M
Instrument : GCMS 6
Sample Name: M-552.2A-06
Misc Info : Dichloroacetic acid in MtBE
Vial Number: 27



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm
Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)
GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

