



=====

Acq. Operator : SYSTEM	Seq. Line :
Sample Operator : SYSTEM	
Acq. Instrument : LCMS-2	Location : P2-C-08
Injection Date : 10/1/2024 6:19:20 AM	Inj : 2
	Inj Volume : 1.000 µl

Different Inj Volume from Sample Entry! Actual Inj Volume : 0.500 µl

Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\100124\100124 2024-10-01 05-57-36\DNPH.M

Last changed : 12/4/2019 12:27:07 PM by SYSTEM

Analysis Method : C:\Users\Public\Documents\ChemStation\1\Methods\Purity_195-315.M

Last changed : 2/6/2023 2:02:18 PM by SYSTEM

Method Info : OPERATOR: Organic QC lab

General gradient method starting @100% water for 1 min @0.3 ml/min, the 100% Acetonitrile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 µm particle size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

