Print of window 38: Current Chromatogram(s)
Data File : C:\USERS\P...ON\1\DATA\07222201\07212201 2022-07-22 10-18-02\02-KD.D
Sample Name : M-554-DNPH-06

==========

Acq. Operator : SYSTEM Seq. Line : 2
Sample Operator : SYSTEM
Acq. Instrument : LCMS-2 Location : P2-C-10
Injection Date : 7/22/2022 10:35:46 AM Inj : 2
Inj Volume : 1.000 µl
Acq. Method : C:\Users\Public\Documents\ChemStation\1\Data\07222201\07212201 2022-07-22 10-18-02\DNPH.M
Last changed : 12/4/2019 12:27:07 PM by SYSTEM
Analysis Method : C:\USERS\PUBLIC\DOCUMENTS\CHEMSTATION\1\METHODS\Purity_195-315.M
Last changed : 3/23/2022 11:04:54 AM by SYSTEM
Method Info : OPERATOR: Organic QC lab
General gradient method starting @100% water for 1 min @0.3 ml/min,the 100% Acetonitile in 13 min, 100% Acetonitrile until T=16 min. Column Zorbax Eclipse plus C-18, 2.1mmx50mm, 1.8 u particlue size. Agilent 1290 Infinity II LC/MS system, Agilent 6125 quadrupole MS detector.

Sample Info : Formaldehyde-DNPH in AcCN/MeOH 5/95

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)